


January 21st, 2021

Certificate of Conformity (CoC) Certificate of Sterility (CoS)

Integrity Part Number	Lot Number	Quantity (each)	Quantity (case)	Sterilization Work Order Number	Sterilization Date
616-1003G	100720-500-120A	80,000	800	2812521	1.20.2021

Integrity Cleanroom certified that the items listed on this certificate have been processed in accordance with our sterilization specifications that were validated in accordance to ANSI/AAMI/ISO 11135 guidelines at a Sterility Assurance Level (SAL) of 10^{-6} .

All records pertaining to sterilization are on file at Integrity Cleanroom.



Mr. Ron Tosh
Chief Operating Officer

Head office
 Integrity House
 Easlea Road
 Bury St Edmunds, IP32 7BY, UK
 T: +44 (0)1473 836 200
 E: info@integritycleanroom.co.uk
 W: www.integritycleanroom.co.uk

North America office
 1000 Heritage Center Circle
 Round Rock
 Texas 78664, U.S.
 T: +1 (0-)512-580-4220
 E: info@integritycleanroom.com
 W: www.integritycleanroom.com

