



Certificate of Processing

STERIGENICS 344 Bonnie Circle Corona CA 92878
 TEL 951 340-0700 FAX www.sterigenics.com

R55480102

12/28/2020 17:49:49 GMT
 Page 1 of 1

Customer Name: Antistat Inc TA Integrity Cleanrooms Processing Facility: Corona Work Order # 2799303
 P.O.# 168942 Sales Order # 2662390
 30-60 kGy INTC_30-60, Gamma Treatment Irradiation Date/Time: 12/26/2020 10:09:00 GMT
 Irradiation Cell: CELL A

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	750	CA	616-1003G		090620-500-120A	N/A
			Dose Map 104_20644	DOSE MAP REQUIRED 2X		
102.000	44	CA	616-1004G		091520-500-120A	N/A
			Dose Map 104_20644	DOSE MAP REQUIRED 2X		
	794	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	30.0 kGy	60.0 kGy	33.9 KGY	Pass	RTOVAR	12/27/2020 03:44:00 GMT
		Reason Code Test				RUBEN TOVAR	
450.00	Maximum Dose	30.0 kGy	60.0 kGy	41.7 KGY	Pass	RTOVAR	12/27/2020 03:44:14 GMT
		Reason Code Test				RUBEN TOVAR	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: ALEX CORDOVA
 Reason: Work Order Completions

Date: 12/28/2020 17:46:33 GMT