



Certificate of Processing

STERIGENICS 3125 Wichita Ct. Fort Worth TX 76140
 TEL 817 293-0999 FAX 817 293-2933 www.sterigenics.com

R55480102

01/21/2021 15:56:31 GMT
 Page 1 of 1

Customer Name: Antistat Inc TA Integrity Cleanrooms Processing Facility: Fort Worth Work Order # 2812521
 P.O.# POR102141 Sales Order # 2679539
 15 - 60 kGy IAT115/60, GAMMA TREATMENT Irradiation Date/Time: 01/20/2021 04:37:00 GMT
 Irradiation Cell: CELL A

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	800	CA	616_1003G	ISOLATION GOWNS	100720-500-120A	NA
	800	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	15.0 kGy	60.0 kGy	34.0 KGY	Pass	JRIDLEY Jake Ridley	01/21/2021 15:20:25 GMT
450.00	Maximum Dose	15.0 kGy	60.0 kGy	43.8 KGY	Pass	JRIDLEY Jake Ridley	01/21/2021 15:20:33 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: MICHAEL BUTLER
 Reason: Work Order Completions

Date: 01/21/2021 15:44:38 GMT